





Configuration proposal General Surgery (CH3)



Configuration description:
 MediBoom® 1000/1000 with Service Head 600
 3 x Shelf 520 mm, 1 x Drawer

1. Brake system: Electromagnetic
2. Action space*: >4,800 mm
3. Weights and Pay Loads
 Remaining Net Pay Load: 109 kg
 Max. pendant pay load: 170 kg
 Vertical force as configured: 2986 N
4. Remaining Diameter: 41 %


* reference: centre of main bearing to centre of shelf

Number of service components											
O ₂	Air 4/5	N ₂ O	Vac	CO ₂	N ₂	AGSS	Air 8/ Air- moto	Electrical sockets	Earth bonding sockets	Double RJ45	Data prep.
1	2		2				1	6	6	1	2
Phoenix P3 British Standard (ISO)								Wands- worth			




Configuration proposal General Surgery (CH3)


Side view right



Front view



Side view left



Back

