




Configuration proposal General Surgery Basic CHST

Configuration description:

OndaScope400 with Service Head 400
1 x shelf 520 mm

1. Brake system: friktion
2. Action space*: ---
3. Weights and pay loads
remaining net pay load: 226 kg
max. pendant pay load: 250 kg
vertical force as configured: 3317 N
4. Remaining diameter: 75%


* reference: centre of main bearing to centre of shelf






Configuration proposal General Surgery Basic CHST


Side view right




Front view



Side view left



Back



Number of service components											
O ₂	Air 4/5	N ₂ O	Vac	CO ₂	N ₂	AGSS	Air 8/ Air- motor	Electrical sockets	Earth bonding sockets	Double RJ45	Dataprep.
1	1		1					4	4		2
Phoenix P3 British Standard (ISO)								Wands- worth			